

Manual Supplement

Manual Title: 233 Getting Started
Part Number: 3465366
Print Date: September 2009
Revision/Date:

Supplement Issue: **2**
Issue Date: 5/10
Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 2, under **Safety Information** add the following:

- CMIID: 2009DJ6034
-  R206WWJUX000013
-  R206WWJUX000012

- 
Agência Nacional de Telecomunicações
0749-10-5988

(01)07898941652

Este equipamento trabalha em caráter secundário, isto é, não tem direito a proteção contra a interferência prejudicial, mesmo de estações do mesmo tipo, e não pode causar interferência a sistemas operando em caráter primário.

Change #2

On page 2, under Warnings, add the following:

- **Remove batteries when not in use for two or more months, this reduces the chance of battery leakage and corrosion.**
- **When your Fluke 233 Digital Multimeter batteries go dead, remove them immediately and replace them with new batteries.**